



TM

Ref. Certif. No.

JPTUV-142316

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME****CB TEST CERTIFICATE**

Product

Centrifuge

Name and address of the applicant

DLAB SCIENTIFIC CO.,LTD.
Yu An Road 31,
Airport Economic core zone, Shun yi District, Beijing,
101318, P.R. China

Name and address of the manufacturer

DLAB SCIENTIFIC CO.,LTD.
Yu An Road 31,
Airport Economic core zone, Shun yi District, Beijing,
101318, P.R. China

Name and address of the factory

DLAB SCIENTIFIC CO.,LTD.
Yu An Road 31,
Airport Economic core zone, Shun yi District, Beijing,
101318, P.R. China

Ratings and principal characteristics

AC 100-120 V, 50/60 Hz, 400 W, Class I for NA
AC 200-240 V, 50/60 Hz, 400 W, Class I for EU

Trademark (if any)

DLAB

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

DM0636, SCI636, 90X5001123
(X=0-9, X stands for different client)Additional information (if necessary may
also be reported on page 2)

For model difference, refer to the test report.

A sample of the product was tested and
found to be in conformity withIEC 61010-1:2010+A1
See Test Report for National DifferencesAs shown in the Test Report Ref. No. which
forms part of this Certificate

CN22T0EQ 001

This CB Test Certificate is issued by the National Certification Body

**TÜVRheinland®**TÜV Rheinland Japan Ltd.
Global Technology Assessment Center
4-25-2 Kita-Yamata, Tsuzuki-ku
Yokohama 224-0021, Japan
Phone + 81 45 914-3888
Fax + 81 45 914-3354
Mail: info@jpn.tuv.com
Web : www.tuv.com

Date: 2022-12-09

Signature:

Miao Mai